

Instrument: ACQ-BSM#H09UPB121M

Binary Solvent Manager Dynamic Leak Test

Accumulator:

Result: Passed

Leak Rate: 108.8 nL/min

Maximum Pressure: 14146 psi

Percent of Final Stroke: 27 %

Compressed Volume: 16 uL

Compression Strokes: 1

Primary:

Result: Passed

Leak Rate: 119.0 nL/min

Maximum Pressure: 12017 psi

Percent of Final Stroke: 25 %

Compressed Volume: 15 uL

Compression Strokes: 1

Test Attempts: 1

